

# CUBE 30 touch

SED-RATE SYSTEM



## TECHNICAL CHARACTERISTICS

**POWER SUPPLY**  
110 to 230 VAC (50 – 60Hz)

**MIXING OF SAMPLES**  
in compliance with CLSI and ICSH requirements

**ABSORBED ELECTRICAL POWER**  
max 100 W

**DIMENSIONS**  
450 x 390 x 297 mm (w x h x d)

**DISPLAY**  
10,1" wide vertical type

**WEIGHT**  
14 Kg

**NOISE LEVEL**  
< 80 Db

**WORKING TEMPERATURE**  
+ 15° to + 35°C

**BAR-CODE READER**  
Automatic internal

**RELATIVE HUMIDITY**  
from 20% to 80% without condensation

## INSTRUMENTS AND ACCESSORIES

CUBE 30 TOUCH	1	Ref. 10395
TEST DEVICE NEXT 500	500 tests	Ref. 10293
TEST DEVICE NEXT 1K	1.000 tests	Ref. 10294
TEST DEVICE NEXT 5K	5.000 tests	Ref. 10296
TEST DEVICE NEXT 10K	10.000 tests	Ref. 10297
ESR CONTROL CUBE	4 x 9 ml	Ref. 10435
ESR CONTROL CUBE	2 x 9 ml	Ref. 10436
EXTERNAL BAR-CODE READER	1	Ref. 20550510
THERMAL PAPER	1	Ref. 10403



**DIESE**  
**DIESSE**

DIESE Diagnostica Senese SpA  
Via del Pozzo, 5 - Loc. San Martino  
53035 Monteriggioni (SI) Italy  
Ph: +39 0577 319560/61 - Fax: +39 0577 318763 - salesoffice@diesse.it  
www.diesse.it

# CUBE 30 touch

SED-RATE SYSTEM



**30 ESR RESULTS  
IN 25 MIN WITH MODIFIED  
WESTERGREN METHOD**

**DIESE**  
**DIESSE**

INNOVATIVE CLINICAL DIAGNOSTIC SYSTEMS

# CUBE130touch

SED-RATE SYSTEM



## TEST PERFORMED ON EDTA/CBC SAMPLES

The test is performed on blood samples collected in the same top lavender tubes used for full blood count.

## USER FRIENDLY

- Up to 30 samples simultaneously
- Internal bar-code reader
- Ease of use
- Reduction of turn-around time
- Standardization
- Internal mixing of samples by complete inversion of the tubes
- 10 inches colour touch screen
- Innovative graphic interface
- 3 working modes: batch, continuous, mixed
- Fast Turn-Around Time (TAT)
- 30 results in less than 25 minutes

## FEATURES

BATCH/CONTINUOUS WORKING MODE


TOUCH SCREEN INTERFACE

REDUCED SIZE

NO SAMPLE CONSUMPTION

NO WASTE FLUIDS

SAFETY



## ENVIRONMENTALLY FRIENDLY

No production of waste materials, without extra-cost for liquid waste disposal.



## ANALYZE DIFFERENT AND NON-DEDICATED TUBES